

# **Exhibit A-6**

# Cleanroom Certification Report

*Service Site:*

New England Compounding Center  
697 Waverly Street  
Framingham, MA 01702

*Prepared For:*

Liberty Industries  
133 Commerce Street  
East Berlin, CT 06023

*Attention:*

Jeff Erickson  
860/828-2653

*Date Tested*

**October 7, 2005**

*Field Service Technician:*


**Charles Kuchinsky**



450 West John Street  
Hicksville, NY 11801  
(516) 433-3700  
FAX (516) 433-9262

Overview

## ANTE ROOM




Ceiling Height = 7.75'

<b>ENV Services, Inc.</b>		<i>Service Specialists for Calibration, Certification and Decontamination</i>		<i>Project: New England Compounding Center Framingham, MA 01702</i>	
450 West John Street		(516) 433-3700		<i>Test Date:</i>	
Hicksville, NY 11801				<b>October 7, 2005</b>	
<i>Technician:</i> <b>CK</b>		<i>Prepared by:</i> deg		<i>Approval:</i> CENV200510025	

ENV-V-Overview-Ante Room

CONFIDENTIAL PROTECTED 001282

HEPA Filter Leak Test  
 ANTE ROOM



OK = No Leakage Detected

ⓧ = Leakage Detected and Repaired

X = Leakage Detected and Not Repaired

<b>ENV Services, Inc.</b> <i>Service Specialists for Calibration, Certification and Decontamination</i>		<b>Project:</b> New England Compounding Center Framingham, MA 01702
450 West John Street Hicksville, NY 11801	(516) 433-3700	<b>Test Date:</b> October 7, 2005
<b>Technician:</b> CK	<b>Prepared by:</b> deg	<b>Approval:</b> CENV200510025


ENV-V-HEPA Leak-Ante Room

CONFIDENTIAL PROTECTED 001283

Velocities

(Measurements in feet per minute)

ANTE ROOM



2 = No. of Readings  
 2% = Maximum variance  
 3% = Relative Std. Deviation  
 99 = Average velocity

<b>ENV Services, Inc.</b> <i>Service Specialists for Calibration, Certification and Decontamination</i>		<b>Project:</b> New England Compounding Center Framingham, MA 01702
450 West John Street Hicksville, NY 11801	(516) 433-3700	<b>Test Date:</b> October 7, 2005
<b>Technician:</b> CK	<b>Prepared by:</b> deg	<b>Approval:</b> CENV200510025


ENV-V-Velocities-Ante Room

CONFIDENTIAL PROTECTED 001284

**Room Pressurization**

(Measurements in inches water gauge)

ANTE ROOM



<b>ENV Services, Inc.</b>		<i>Service Specialists for Calibration, Certification and Decontamination</i>		<i>Project: New England Compounding Center Framingham, MA 01702</i>	
450 West John Street		(516) 433-3700		<i>Test Date:</i>	
Hicksville, NY 11801				October 7, 2005	
<i>Technician: CK</i>		<i>Prepared by: deg</i>		<i>Approval: CENV200510025</i>	


ENV-V-Pressurization-Ante Room

CONFIDENTIAL PROTECTED 001285

**Particle Counts**

(Per cubic foot, measured at 0.5 microns and larger)

ANTE ROOM




	<b>Lgth</b>	<b>Wdth</b>									
<b>Dim.</b>	<b>10</b>	<b>7.7</b>	<b>AT .5 MICRONS AND LARGER</b>								
<b>Area(ft<sup>2</sup>)</b>	<b>77</b>		<b># of</b>	<b>High Location</b>		<b>Avg. of Loc.</b>		<b>95% UCL</b>		<b>FS209E</b>	<b>ISO</b>
<b>Area(m<sup>2</sup>)</b>	<b>7.15</b>		<b>Locations</b>	<b>ft<sup>3</sup></b>	<b>m<sup>3</sup></b>	<b>ft<sup>3</sup></b>	<b>m<sup>3</sup></b>	<b>ft<sup>3</sup>*</b>	<b>m<sup>3</sup></b>	<b>Class</b>	<b>Class</b>
			5	4385	154,855	2353	83,081	3439	120,888	100,000	8
<b>ISO</b>			* ISO Class 8 allows a maximum of 3,520,000 particles 0.5 microns and larger per cubic meter as an average count at any location as long as the entire clean area meets the required statistical tests.								
<b>Target</b>	<b>8</b>		-----								
<b>Class</b>			* For comparison the calculated values are presented in both cubic feet and cubic meters.								
			* The 95%UCL calculated for cubic feet uses the FS209E UCL Factors.								

<b>ENV Services, Inc.</b>				<b>Project New England Compounding Center</b>			
<i>Service Specialists for Calibration, Certification and Decontamination</i>				<b>Framingham, MA 01702</b>			
<b>450 West John Street</b>		<b>(516) 433-3700</b>		<b>Test Date:</b>			
<b>Hicksville, NY 11801</b>				<b>October 7, 2005</b>			
<b>Technician: CK</b>		<b>Prepared by: deg</b>		<b>Approval: CENV200510025</b>			

ENV-V-.5 ISO-Ante Room

CONFIDENTIAL PROTECTED 001286

Overview

Ceiling Height = 7.75'

<b>ENV Services, Inc.</b>		<i>Service Specialists for Calibration, Certification and Decontamination</i>		<b>Project:</b> New England Compounding Center Framingham, MA 01702	
450 West John Street		(516) 433-3700		<b>Test Date:</b>	
Hicksville, NY 11801				October 7, 2005	
<b>Technician:</b> CK		<i>Prepared by:</i> deg		<i>Approval:</i> GENV200510025	


ENV-V-Overview-Cleanroom

CONFIDENTIAL PROTECTED 001287



**HEPA Filter Leak Test**

## CLEANROOM



OK = No Leakage Detected

⊗ = Leakage Detected and Repaired

X = Leakage Detected and Not Repaired


<b>ENV Services, Inc.</b>		Project: New England Compounding Center	
Service Specialists for Calibration, Certification and Decontamination		Framingham, MA 01702	
450 West John Street		Test Date:	
(516) 433-3700		October 7, 2005	
Hicksville, NY 11801		Approval: CENV200510025	
Technician: CK	Prepared by: deg		

ENV-V-HEPA Leak-Cleanroom

CONFIDENTIAL PROTECTED 001288

Velocities

(Measurements in feet per minute)  
CLEANROOM



14 = No. of Readings  
10% = Maximum variance  
5% = Relative Std. Deviation  
97 = Average velocity

<b>ENV Services, Inc.</b> <i>Service Specialists for Calibration, Certification and Decontamination</i>		<b>Project:</b> New England Compounding Center Framingham, MA 01702
450 West John Street Hicksville, NY 11801	(516) 433-3700	<b>Test Date:</b> October 7, 2005
<b>Technician:</b> CK	<b>Prepared by:</b> deg	<b>Approval:</b> CENV200510025


ENV-V-Velocities-Cleanroom

CONFIDENTIAL PROTECTED 001289

Room Pressurization

(Measurements in inches water gauge)

CLEANROOM




ENV Services, Inc. <i>Service Specialists for Calibration, Certification and Decontamination</i>		Project: New England Compounding Center Framingham, MA 01702
450 West John Street Hicksville, NY 11801	(516) 433-3700	Test Date: October 7, 2005
Technician: CK	Prepared by: deg	Approval: CENV200510025

ENV-V-Pressurization-Cleanroom

CONFIDENTIAL PROTECTED 001290

**Particle Counts**

(Per cubic foot, measured at 0.5 microns and larger)

**CLEANROOM**

Area(ft<sup>2</sup>) 478.8  
 Area(m<sup>2</sup>) 44.48

ISO  
 Target 7  
 Glass

**AT .5 MICRONS AND LARGER**

# of Locations	High Location		Avg. of Loc.		95% UCL		FS209E Class	ISO Class
	ft <sup>3</sup>	m <sup>3</sup>	ft <sup>3</sup>	m <sup>3</sup>	ft <sup>3</sup> *	m <sup>3</sup>		
14	1726	60,953	739	26,085	N/A	N/A	10,000	7

\* ISO Class 7 allows a maximum of 352,000 particles 0.5 microns and larger per cubic meter as an average count at any location as long as the entire clean area meets the required statistical tests.

-----

\* For comparison the calculated values are presented in both cubic feet and cubic meters.

\* The 95%UCL calculated for cubic feet uses the FS209E UCL Factors.

**ENV Services, Inc.**

Service Specialists for Calibration, Certification and Decontamination

450 West John Street  
Hicksville, NY 11801

(516) 433-3700

Project: New England Compounding Center  
Framingham, MA 01702Test Date:  
October 7, 2005

Technician: CK

Prepared by: deg

Approval: CENV200510025

ENV-V-.5 ISO-Cleanroom

CONFIDENTIAL PROTECTED 001291

Test Equipment**Primary Tests**Photometer

Manufacturer: ATI  
 Model No. TDA-2G  
 Serial No. 11883  
 Calibration Date: 3/22/2005

Particle Counter

Manufacturer: MET ONE  
 Model No. A2400  
 Serial No. 040501029  
 Calibration Date: 5/17/2005

AirData Multimeter

Manufacturer: Shortridge  
 Model No. ADM-870  
 Serial No. M90364  
 Calibration Date: 6/14/2005

<b>ENV Services, Inc.</b> <i>Service Specialists for Calibration, Certification and Decontamination</i>		Project: <b>New England Compounding Center</b> <b>Framingham, MA 01702</b>
450 West John Street Hicksville, NY 11801	(516) 433-3700	Test Date: <b>October 7, 2005</b>
Technician: <b>CK</b>	Prepared by: <b>deg</b>	Approval: <b>CENV200510025</b>

ENV-V-Test Equipment

CONFIDENTIAL PROTECTED 001292



# TEC Services, Inc.

90 Aileron Court, Suite 9  
Westminster, Maryland 21157  
Phone: 410-871-0573  
Fax: 410-871-0574

Controlled Environment Products  
Calibration & Services

## EQUIPMENT CALIBRATION REPORT

Date Received: 3/22/05 Date Returned: 3/22/05  
Customer: ENV Services Inc. Contact: Aurora  
Address: 450 West John Street Phone #: 516-433-3700  
Hicksville, NY 11801 Customer-PO #: 2575  
Manufacturer: Air Techniques Model #: TDA-2G  
Instrument Type: Aerosol Photometer Serial #: 11883

### INSTRUMENT DATA

		As Found	Final	Mfg. Tolerances		As Found	Final	Mfg. Tolerances
Straylight	(3)	.0011%	.0009%	Not Applicable	U4-3	(1)	-11.98	-11.98
Sample Flow	(1)	29.0	29.0	28.3 ± 2.8 SLPM	U8-1	(1)	12.04	12.04
0.001%	(1)	.80 x 10 <sup>-10</sup>	.80 x 10 <sup>-10</sup>	.80 ± .04 x 10 <sup>-10</sup>	U12-6	(1)	5.00	5.00
0.01%	(1)	.80 x 10 <sup>-9</sup>	.80 x 10 <sup>-9</sup>	.80 ± .04 x 10 <sup>-9</sup>	U13-1	(1)	9.72	9.72
0.1%	(1)	.80 x 10 <sup>-8</sup>	.80 x 10 <sup>-8</sup>	.80 ± .04 x 10 <sup>-8</sup>	J6-Wh	(1)	295	295
1.0%	(1)	.80 x 10 <sup>-7</sup>	.80 x 10 <sup>-7</sup>	.80 ± .04 x 10 <sup>-7</sup>	J9-1	(1)	5.07	5.07
10%	(1)	.80 x 10 <sup>-6</sup>	.80 x 10 <sup>-6</sup>	.80 ± .04 x 10 <sup>-6</sup>	J9-5	(1)	14.99	14.99
100%	(1)	.80 x 10 <sup>-5</sup>	.80 x 10 <sup>-5</sup>	.80 ± .04 x 10 <sup>-5</sup>	J9-6	(1)	-14.97	-14.97

(1) In Tolerance (2) Out of Tolerance (3) No effect on instrument operation (N/A) Not Applicable \* Amperes

Received in Tolerance: ☒ Yes ☐ No ☐ Not Operable ☐ Tolerance Not Applicable

### MAINTENANCE PERFORMED

- |   |  |   |
|---|--|---|
| <input checked="" type="checkbox"/> Clean / Rebuild Chamber | <input type="checkbox"/> Replace Cell Lamp   | <input checked="" type="checkbox"/> Leak Test Sampling System |
| <input checked="" type="checkbox"/> Flush Hoses & Pump      | <input type="checkbox"/> Test Scanning Probe | <input checked="" type="checkbox"/> Perform Voltage Check     |
| <input checked="" type="checkbox"/> Check Wiring & Hardware | <input type="checkbox"/> Replace Gaskets     | <input checked="" type="checkbox"/> Realign Optics            |
| <input type="checkbox"/>                                    |  | <input checked="" type="checkbox"/> Performance Test          |

Calibrated for: ☒ PAO ☒ DOP

Temperature: 22.0 °C Relative Humidity: 36.0 %

Notes: Installed 2 bumper feet. Received without probe, cord, and 1 side panel.


### NIST TRACEABLE EQUIPMENT USED

Description	Model	Serial #	Certificate #	Cal Date	Due Date
Picoampere Source	Keithley 261	72255	5-6801	05 APR 04	05 APR 05
Digital Multimeter	Fluke 87-3	78950234	1000233119	07 FEB 05	07 FEB 06
Thermohygrometer	Oakton WD-35700-00	N/A	1000233120	04 FEB 05	04 FEB 06
Flowmeter	Dwyer VFB-69-BV	N/A	1000233121	09 FEB 05	09 FEB 06

Calibrated per Procedure: ☒ TEC001 ☒ TEC002 ☐ TEC003 ☐ TEC004 ☐ TEC005 ☒ TEC006  
☒ TEC007 ☒ TEC008 ☒ TEC009 ☐ TEC010 ☐

CALIBRATION BY: CRP DATE: 3/22/05 DUE DATE: 3/22/06  
FINAL TEST & APPROVAL BY: T.E.

Certificate No. : 62911  
 Procedure No. : SCP-76.1



Control No. : 62911  
 Page 1 of 1

**UNIT UNDER TEST**

Manufacturer : MET ONE  
 Model No. : A2400  
 Serial No. : 040501029  
 Cust. Ref. No. :  
 Description : PARTICLE COUNTER  
 Date Rec'd : 05/13/2005  
 Condition Rec'd : Good

**SUBMITTED BY**

Customer: ENV SERVICES, INC.(N.E.)  
 450 W JOHN STREET  
 HICKSVILLE, NY 11801  
 P.O.# : 033813  
 Precal: OUT OF SPEC Final : IN SPEC

**CALIBRATION CERTIFICATE**

All calibrations are performed by qualified personnel using instrumentation, procedures and methods which guarantee specifications claimed are reliable. When specified, all calibrations are performed in accordance with ISO/IEC 17025, ANSI/NCCL Z-540-1-1994, MIL-STD-45662A, 10CFR21 and Protronics/Pro-Lab Quality Manual (Per current revision). Standards used are traceable to The National Institute of Standards and Technology (NIST). Expanded uncertainties are calculated using methods described in the Guide to the Expression of Uncertainty of Measurement (GUM) utilizing a coverage factor of K=2 (95% confidence) and kept on file at Pro-Lab. At a minimum, standards are selected with an uncertainty of 25% or better, where possible. This certificate and/or data shall not be reproduced except in full, without the written permission of Pro-Lab management.

**Standards Used**

Asset#	Description	Certificate Number	Measurement Function	Measurement Uncertainty	Date Due
132	<u>TIMER, TRANSFER STANDARD</u>	58333			11/4/2005
170	<u>PARTICLE COUNTER</u>	59311			11/24/2005
173	<u>MASS FLOWMETER - 2SCFM</u>	56310			6/28/2005

Temperature : 24.0 °C  
 Humidity : 35.0 % RH

Date Tested : 05/17/2005  
 Date Due : 05/17/2006

Approved By: Michael Parente, Jr.

Laboratory Manager


Calibrated By: Mike Blahut

Calibration Technician

— AN ENV COMPANY —

2880 Bergey Rd. - Suite K - Hatfield, PA 19440-1764 - 800-992-9108 - (FAX) 215-822-6522

Certificate No. : 63016  
 Procedure No. : MFGR



Control No. : 63016  
 Page 1 of 1

**UNIT UNDER TEST**

Manufacturer : SHORTRIDGE  
 Model No. : ADM-870  
 Serial No. : M90364  
 Cust. Ref. No. :  
 Description : AIR DATA MULTIMETER  
 Date Rec'd : 05/19/2005  
 Condition Rec'd : Fair

**SUBMITTED BY**

Customer: ENV SERVICES, INC.(N.E.)  
 450 W JOHN STREET  
 HICKSVILLE, NY 11801  
 P.O.# :  
 Precal: IN SPEC Final : IN SPEC

**CALIBRATION CERTIFICATE**

All calibrations are performed by qualified personnel using instrumentation, procedures and methods which guarantee specifications claimed are reliable. When specified, all calibrations are performed in accordance with ISO/IEC 17025, ANSI/NCSL Z-540-1-1994, MIL-STD-45662A, 10CFR21 and Protronics/Pro-Lab Quality Manual (Per current revision). Standards used are traceable to The National Institute of Standards and Technology (NIST). Expanded uncertainties are calculated using methods described in the Guide to the Expression of Uncertainty of Measurement (GUM) utilizing a coverage factor of K=2 (95% confidence) and kept on file at Pro-Lab. At a minimum, standards are selected with an uncertainty of 25% or better, where possible. This certificate and/or data shall not be reproduced except in full, without the written permission of Pro-Lab management.

**Standards Used**

#	Description	Certificate Number	Measurement Function	Measurement Uncertainty	Date Due
167	DEAD WEIGHT TESTER	51884			10/22/2005

Temperature : 23.0 °C  
 Humidity : 46.0 % RH

Date Tested : 06/14/2005  
 Date Due : 06/14/2006

Approved By: Michael Parente, Jr.

Laboratory Manager

Calibrated By: Keith R. Murray Sr.

Calibration Technician

— AN ENV COMPANY —

2880 Bergey Rd. - Suite K - Hatfield, PA 19440-1764 - 800-992-9108 - (FAX) 215-822-6522





# ***CERTIFICATE OF COMPLIANCE***

## ***FOR***

**New England Compounding Center  
Framingham, MA 01702**

### **Ante Room**

The above referenced clean zone has successfully passed the minimum requirements for an ISO Class 8 clean zone per the requirements of ISO Standards 14644-1:1999 and 14644-2:2000 for particle count testing of Cleanrooms and associated controlled environments.

This clean zone was tested in the At Rest occupancy state at 0.5 micrometers and larger particle sizes. Testing was performed as outlined in ISO Standards 14644-1:1999 and 14644-2:2000. The classification is applicable to the above clean zone designated in the report prepared October-2005.

Date of Certification: October 7, 2005

Due for Retest: April-2006

Certified By: Charles Kuchinsky

Approval: CENV200510025



# ***CERTIFICATE OF COMPLIANCE***

## ***FOR***

**New England Compounding Center  
Framingham, MA 01702**

### **Cleanroom**

The above referenced clean zone has successfully passed the minimum requirements for an ISO Class 7 clean zone per the requirements of ISO Standards 14644-1:1999 and 14644-2:2000 for particle count testing of Cleanrooms and associated controlled environments.

This clean zone was tested in the At Rest occupancy state at 0.5 micrometers and larger particle sizes. Testing was performed as outlined in ISO Standards 14644-1:1999 and 14644-2:2000. The classification is applicable to the above clean zone designated in the report prepared October-2005.

Date of Certification: October 7, 2005

Due for Retest: April-2006

Certified By: Charles Kuchinsky

Approval: CENV200510025